



## EUA Language

EUA Language that must appear on any materials for the Covid-19/ Flu A&B Combo Rapid Test Cassette (Swab) product, including but not limited to web site, social media, marketing materials, customer letters, etc.

### Authorized Laboratories:

**Laboratories certified under the Clinical Laboratory Improvement Amendments of 1988 (CLIA), 42 U.S.C. §263a, that meet the requirements to perform moderate, high, or waived complexity tests. This product is authorized for use at the Point of Care (POC), i.e., in patient care settings operating under a CLIA Certificate of Waiver, Certificate of Compliance, or Certificate of Accreditation.**

The Healgen® COVID-19/Flu A&B Ag Combo Rapid test Cassette (Swab) has not been FDA cleared or approved but has been authorized by FDA under an EUA for use by authorized laboratories.

This product has been authorized only for the detection of proteins from SARS-CoV-2, influenza A and influenza B, not for any other viruses or pathogens.

The emergency use of this product is only authorized for the duration of the declaration that circumstances exist justifying the authorization of emergency use of in vitro diagnostics for detection and/or diagnosis of COVID-19 under Section 564(b)(1) of the Federal Food, Drug and Cosmetic Act, 21 U.S.C. § 360bbb-3(b)(1), unless the declaration is terminated or authorization is revoked sooner.